

ES certifikat o pregledu zasnove
EC Design Examination Certificate

Pregled zasnove merila v skladu s Pravilnikom o merilnih instrumentih
Design examination according to EC Directive on Measuring Instruments

Št. / No.: 08MID002
Popravek / Revision: R02

Ime in naslov proizvajalca / Name and address of the manufacturer :

Proizvajalec / *Manufacturer* : ISKRAEMECO, merjenje in upravljanje energije, d.d.
Naslov / *Address* : Savska loka 4, 4000 Kranj, SLOVENIJA

Podatki o merilu / Description of the measuring instrument :

Merilo / *Measuring instrument* : Trifazni elektronski večfunkcijski števec električne energije /
Three-phase multi-function electronic meter
Tip / *Type* : MT830, MT831
Referenčne napetosti /
Reference voltages : 3x230/400V, 3x240/415V, 3x57,7/100V ... 3x240/415V
Referenčni tokovi /
Reference Currents : 5 A, 10 A
Nazivni tokovi /
Nominal Currents : 1 A, 5 A
Razred točnosti /
Accuracy Class : A, B, C

V skladu s Pravilnikom o merilnih instrumentih (UL RS št. 42/2006), poglavje MI-003 – Števci delovne električne energije, dodatek H1, člen 4, je naročnik predložil vlogo, tehnično dokumentacijo in dokazila o ustreznosti zasnove v pregled zasnove merila za zgoraj navedeni proizvod, z namenom, da se preveri ali zasnova proizvoda ustreza zahtevam tega pravilnika. / *In accordance with the Directive on Measuring Instruments 2004/22/EC, Annex MI-003 – Active Electrical Energy Meters, Annex H1, article 4, the applicant has submitted the application, technical documentation and the supporting evidence for the adequacy of the technical design for the above mentioned measuring instrument for the purpose of design examination. This is to certify, that the design of the measuring instrument meets the provisions laid down in the Directive.*

V skladu s Pravilnikom o merilnih instrumentih mora naročnik obvestiti priglašeni organ o vsaki narejeni ali načrtovani spremembi. / *In accordance with the above mentioned Directive the applicant has to inform the notified body of any already performed or planned modifications.*

Pregledana tehnična mapa se shrani pri priglašenem organu za dobo 10 let po izdelavi zadnjega primerka merila. Na željo naročnika se mapa predmeta po tem obdobju vrne naročniku ali uniči. / *The examined technical file will be stored by the notified body for 10 years after the last measuring instrument has been manufactured. On request of the applicant, it will then be returned or destroyed.*

Opomba / Remark:

Ta ES certifikat o pregledu zasnove je popravek certifikata št. 08MID002 popravek R01. /
This EC Design Examination Certificate corrects Certificate No. 08MID002 revision R01.

Certifikat ima prilogo, ki vsebuje 38 strani. /
Certificate has an Annex, which includes 38 pages.

Ljubljana, 2008-10-14

SIQ
Slovenski institut za
kakovost in meroslovje
Slovenian Institute of
Quality and Metrology
Tržaška 2, 1000 Ljubljana

Podpis pooblaščenice osebe / *Authorised signature*

Alja Pregl